

APR 11 2006

Page 1 of 2

Section II

Summary of Safety and Effectiveness (as required by 21 CFR 807.92)

K060680

Atrilaze™ Surgical Ablation System

Submitter: MedicalCV, Inc.
9725 South Robert Trail
Inver Grove Heights, MN 55077
USA

Contact: Denny Steger
V.P. RA/QA
Phone: 651 452 3000
Fax: 651 452 4948

Date of Summary: March 14, 2006

Classification Name: Laser Instrument,
Surgical Powered

Common Name: Surgical Laser Instrument

Proprietary Atrilaze™ Surgical
Ablation System

Description of Device: The Atrilaze Surgical Ablation System consists of a generator designed for the delivery of 810nm or 1064nm laser light and a hand held fiber optic light delivery device (probe) fitted with a standard SMA 905 connector at the proximal end. The system may be used in conjunction with surgical treatment for hemostasis, incision, ablation, coagulation and vaporization of tissue as required by the clinician.

Statement of Intended Use: The MedicalCV Atrilaze Surgical Ablation System is indicated for the delivery of 810nm or 1064nm laser light to soft tissue to include cardiac tissue, during surgical procedures. Indications include the incision, excision, dissection, vaporization, ablation or coagulation of soft tissue.

Warning: The Atrilaze™ Surgical Ablation System is not indicated for the treatment of cardiac arrhythmias.

The risk of actual damage to adjacent organs from the instrument exists and perforation, rupture or tearing of tissue, may occur as a complication of laser use. Burns can occur if the laser energy is not correctly applied. These complications may be serious.

Technological Comparison: The current Atrilaze Surgical Ablation System (K040744 & K052495) consists of a standalone surgical laser emitting 810nm light and a disposable fiber optic delivery system (probe). The fiber optic delivery system is coupled to the laser via an SMA 905 connector to deliver laser radiation to the targeted tissue. The proposed modification to the currently marketed Atrilaze™ Surgical Ablation System is to provide the end-user with an alternate laser generator (wavelength) for ablating soft tissue by providing the end user the option of using either a gallium aluminum arsenide (GaAlAs) semiconductor diode laser at a wavelength of 810nm or a Neodymium-yttrium-aluminum-garnet (Nd:YAG) laser at a wavelength of 1064nm. There is no change to the currently available fiber optic delivery systems (probes), which are coupled to either laser via an SMA 905 connector to deliver laser radiation to the target tissue. For purposes of this submission, the Nd:YAG laser (1064nm) used in conjunction with the Atrilaze Surgical Ablation System was compared to the following predicate device:

- Medical CV Atrilaze™ Surgical Ablation System (K040744 & K052495)

Testing: The results of biocompatibility testing conducted on the disposable fiber optic delivery systems (probes) contained in the original submissions (K040744 & K052495) support that the materials used in the manufacture of the disposable probe are non-toxic, non-hemolytic, and non-pyrogenic. All biocompatibility testing was conducted under Good Laboratory Practices per 21 CFR Part 58. Since there has been no change to the design, packaging and/or sterilization process for the disposable probe additional biocompatibility or sterilization testing was determined not to be required.

Performance testing for the Atrilaze Surgical Ablation System utilizing the Nd:YAG 1064nm laser included compliance to manufacturing specifications for Power Output, with visual and histology evaluation of the lesions obtained using 1064nm wavelength on cardiac tissue. Testing demonstrated that lesions obtained using the Nd:YAG laser at 1064nm wavelength are substantially equivalent to those obtained with the currently cleared laser at 810nm wavelength.



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

APR 11 2006

MedicalCV, Inc.
c/o Mr. Denny Steger
VP, Regulatory Affairs/Quality Assurance
9725 South Robert Trail
Inver Grove Heights, Minnesota 55077

Re: K060680

Trade/Device Name: Atrilaze™ Surgical Ablation System

Regulation Number: 21 CFR 878.4810

Regulation Name: Laser surgical instrument for use in general and plastic surgery
and in dermatology

Regulatory Class: II

Product Code: GEX

Dated: March 14, 2006

Received: March 15, 2006

Dear Mr. Steger:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set


Page 2 – Mr. Denny Steger

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0115. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/dsma/dsmamain.html>

Sincerely yours,

A handwritten signature in black ink, appearing to read "M. Melkerson", with a stylized initial "M" and a trailing flourish.

Mark N. Melkerson
Director
Division of General, Restorative
and Neurological Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use Statement

510(K) Number: K060680

Device Name: Atrilaze™ Surgical Ablation System

Indications for Use: The Medical CV Atrilaze™ Surgical Ablation System is indicated for delivery of 810nm or 1064nm laser light to soft tissue to include cardiac tissue during surgical procedures. Indications include the incision, excision, dissection, vaporization, ablation, or coagulation of soft tissue.

Warning: The Atrilaze™ Surgical Ablation System is not indicated for the treatment of cardiac arrhythmias.

The risk of actual damage to adjacent organs from the instrument exists and perforation, rupture or tearing of tissue, may occur as a complication of laser use. Burns can occur if the laser energy is not correctly applied. These complications may be serious.


Prescription Use X
(Per 21 CFR 801.109)

OR

Over-the-Counter Use _____

(Please do not write below this line -- Continue on another page if necessary)

Concurrence of CDRH, Office of Device Evaluation (ODE)



(Division Sign-Off)

**Division of General, Restorative,
and Neurological Devices**

510(k) Number K060680